

UKCA DECLARATION OF CONFORMITY

Manufacturer: Heathrow Scientific® LLC, 620 Lakeview Parkway, Vernon Hills, IL 60061 USA

Product: Sprout® Plus

Product Model No(s): HS120610, HS120611, HS120612, HS120613, HS120614

This unit has been constructed and conforms to the following:

Directives and Regulations:

Electrical Equipment (Safety) Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

The Electromagnetic Compatibility Regulations 2016

EMC Standards:

EN 61326-1:2013-b (IEC61326-1:2012)

ICES-003, Issue 6

Safety Standards:

EN/IEC 61010-1 2010 ed.3

EN/IEC 61010-2-020 ed.3, 2016

ANSI/UL Std. 61010-1 ed.3, 2012

ANSI/UL Std. 61010-2-020 ed.3, 2016

CAN/CSA C22.2 No.61010-1-12 (R2012)

CAN/CSA C22.2 No.61010-2-020 (R2017)

Signed:



Gary Kamees

Title: Vice President Product Development/Manufacturing

Location: Vernon Hills, IL

Date of Issue: October 2022



Note: Changes or modifications to the product not expressly approved by the manufacturer could void the user's authority to operate the equipment.

Note: This product has been tested and found to comply with the limits of a Class 1A digital device pursuant to Part 15 of the FCC rules. These limits are designed to provide reasonable protection against harmful interference when the product is operating in a commercial environment. The product generates, uses, and can radiate radio frequency energy and if not installed and used in accordance with the user's manual may cause harmful interference to radio communications. Operation of the product in a residential area is likely to cause interference in which case the user will be required to correct the interference.

